
Midmark M11 - v1.0.0
Total Cycles: 4028

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2023 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 147.8 F 1.6
4:00 150.0 F 4.3
6:00 171.0 F 7.8
8:00 192.7 F 11.9
10:00 210.0 F 4.1
12:00 219.1 F 8.4
14:00 231.1 F 12.9
16:00 233.3 F 9.0
18:00 240.5 F 11.0
20:00 244.3 F 12.5
22:00 255.8 F 18.6
24:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.8 F 28.5
3:00 271.5 F 28.6
4:00 272.6 F 29.1

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 216.9 F 0.0
10:00 230.1 F 0.0
15:00 226.2 F 0.0
20:00 227.6 F 0.0
25:00 231.7 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:14
CYCLE COMPLETE

11/16/2023 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____