
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005772
7:56 19/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:55
STERILIZING 4:00
136.4 C 220kPa 4:00
136.3 C 223kPa 4:31
136.3 C 218kPa 5:01
136.3 C 221kPa 5:31
136.4 C 219kPa 6:01
136.3 C 222kPa 6:31
136.2 C 221kPa 7:01
136.4 C 222kPa 7:31
136.4 C 222kPa 7:31
VENTING 7:31
AIR DRYING 8:18
CYCLE COMPLETE 8:26

Digital Signature #
10336DB2EECD332C
