
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000279
12:36 23/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:14
STERILIZING 4:13
136.0 C 220kPa 4:13
136.1 C 223kPa 4:43
136.4 C 223kPa 5:13
136.0 C 221kPa 5:43
136.4 C 224kPa 6:13
136.2 C 221kPa 6:43
136.1 C 220kPa 7:13
136.4 C 225kPa 7:43
136.4 C 225kPa 7:43
VENTING 7:43
AIR DRYING 8:32
CYCLE COMPLETE 8:41

Digital Signature #
B7B777C3964F57A2
