
Midmark M11 - v1.0.5
Total Cycles: 1107

Name: RITTER2
Sterilizer ID: V2313228
05/18/2022 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.5 F 0.0
2:00 125.4 F 0.0
4:00 134.5 F 0.9
6:00 158.6 F 4.4
8:00 192.2 F 8.5
10:00 213.6 F 13.5
12:00 222.2 F 7.8
14:00 234.4 F 12.5
16:00 234.1 F 9.5
18:00 238.8 F 9.8
20:00 245.7 F 13.7
22:00 257.1 F 19.9
24:00 267.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.3 F 29.5
2:00 271.5 F 29.5
3:00 271.9 F 29.4
4:00 271.6 F 29.5

Min 271.2 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 210.3 F 0.0
10:00 211.5 F 0.0
15:00 211.3 F 0.0
20:00 212.7 F 0.0
25:00 215.4 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:16
CYCLE COMPLETE

05/18/2022 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____