
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006437
8:19 11/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:43
STERILIZING 4:48
136.5 C 224kPa 4:48
136.1 C 220kPa 5:18
136.3 C 222kPa 5:48
136.4 C 219kPa 6:18
136.1 C 221kPa 6:48
136.0 C 218kPa 7:18
136.1 C 220kPa 7:48
136.4 C 221kPa 8:18
136.4 C 221kPa 8:18
VENTING 8:18
AIR DRYING 9:11
CYCLE COMPLETE 9:14

Digital Signature #
70BA78435C8667A2
