
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004389
9:21 20/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:36
STERILIZING 2:35
136.0 C 220kPa 2:35
136.3 C 222kPa 3:05
136.3 C 220kPa 3:35
136.2 C 222kPa 4:05
136.3 C 221kPa 4:35
136.1 C 217kPa 5:05
136.4 C 223kPa 5:35
136.4 C 222kPa 6:05
136.4 C 222kPa 6:05
VENTING 6:05
AIR DRYING 6:56
CYCLE COMPLETE 8:58

Digital Signature #
77C047368B7C6A8E
