
Midmark M11 - v1.0.0
Total Cycles: 3207

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2023 09:50 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.4 F 0.0
2:00 179.6 F 2.2
4:00 190.3 F 7.8
6:00 208.7 F 9.1
8:00 230.1 F 9.0
10:00 234.2 F 8.1
12:00 243.2 F 11.4
14:00 246.9 F 13.4
16:00 261.0 F 21.0
18:00 272.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.1 F 28.0
2:00 272.8 F 28.6
3:00 271.7 F 28.1

Min 271.4 F 27.7
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 222.4 F 0.0
Dry Cycle Aborted

FILLING: 0:55
HEATING: 18:18
STERILIZING: 3:00
VENTING: 2:52
DRYING: 1:24
TOTAL CYCLE: 00:26:29

05/05/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____