

-----  
Midmark M11 - v1.0.5  
Total Cycles: 131

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2021 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.5 F 0.0  
2:00 168.0 F 3.0  
4:00 189.8 F 7.5  
6:00 212.0 F 13.2  
8:00 221.8 F 8.4  
10:00 233.8 F 10.3  
12:00 238.7 F 12.2  
14:00 239.6 F 11.0  
16:00 252.6 F 17.8  
18:00 264.2 F 24.7  
20:00 270.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.8  
1:00 273.0 F 30.5  
2:00 271.8 F 29.7  
3:00 273.0 F 30.7  
4:00 272.5 F 30.2

Min 271.6 F 29.7  
Max 273.2 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.2 F 0.0  
5:00 217.9 F 0.0  
10:00 206.2 F 0.0  
15:00 206.9 F 0.0  
20:00 206.2 F 0.0  
25:00 176.6 F 0.0  
30:00 180.3 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 00:59:14  
CYCLE COMPLETE

06/23/2021 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_