
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002249
10:15 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:41
136.4 C 223kPa 3:41
136.4 C 225kPa 4:11
136.0 C 223kPa 4:41
136.3 C 225kPa 5:11
136.2 C 222kPa 5:41
136.4 C 225kPa 6:11
136.3 C 224kPa 6:41
136.2 C 224kPa 7:11
136.2 C 224kPa 7:11
VENTING 7:11
AIR DRYING 8:03
CYCLE COMPLETE 38:03

Digital Signature #
5F7250B55DD7725C
