

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000372  
10:26 21/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:19  
STERILIZING 3:24  
136.0 C 220kPa 3:24  
136.4 C 233kPa 3:54  
136.0 C 231kPa 4:24  
136.4 C 233kPa 4:54  
136.0 C 231kPa 5:24  
136.1 C 229kPa 5:54  
136.2 C 232kPa 6:24  
136.4 C 233kPa 6:54  
136.4 C 233kPa 6:54  
VENTING 6:54  
AIR DRYING 7:43  
CYCLE COMPLETE 7:49

-----  
Digital Signature #  
82CECE69DD2E8664  
-----