

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003781  
14:27 07/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:44  
STERILIZING 3:57  
136.4 C 220kPa 3:57  
136.0 C 222kPa 4:27  
136.0 C 222kPa 4:57  
136.2 C 224kPa 5:27  
136.3 C 223kPa 5:57  
CYCLE ABORTED 6:02  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 6:02  
-----

Digital Signature #  
5B9C861A444997E1  
-----