
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000786
10:48 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:06
STERILIZING 3:06
136.0 C 228kPa 3:06
136.3 C 233kPa 3:36
136.4 C 233kPa 4:06
136.4 C 231kPa 4:36
136.2 C 232kPa 5:06
136.4 C 233kPa 5:36
136.3 C 233kPa 6:06
136.4 C 233kPa 6:36
136.4 C 233kPa 6:36
VENTING 6:36
AIR DRYING 7:28
CYCLE COMPLETE 7:31

Digital Signature #
FC92CD2008318D5C
