
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006041
7:43 01/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:57
STERILIZING 4:00
136.2 C 220kPa 4:00
136.3 C 221kPa 4:30
136.4 C 222kPa 5:00
136.4 C 223kPa 5:30
136.4 C 219kPa 6:00
136.4 C 219kPa 6:30
136.4 C 221kPa 7:00
136.4 C 223kPa 7:30
136.4 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:17
CYCLE COMPLETE 8:19

Digital Signature #
D13498C3022A2BE0
