
Midmark M11 - v1.0.5
Total Cycles: 3540

Name: RITTER2
Sterilizer ID: V2313228
04/18/2024 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 69.8 F 0.0
4:00 80.4 F 0.0
6:00 110.2 F 0.2
8:00 140.0 F 1.9
10:00 163.7 F 4.3
12:00 187.2 F 7.1
14:00 195.4 F 10.9
16:00 212.1 F 6.1
18:00 217.6 F 7.4
20:00 227.4 F 11.4
22:00 228.1 F 7.3
24:00 240.4 F 12.5
26:00 239.4 F 10.8
28:00 251.2 F 16.6
30:00 261.6 F 22.6
32:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 30.0
2:00 271.9 F 29.4
3:00 271.5 F 29.5
4:00 272.7 F 30.0

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 203.5 F 0.0
10:00 204.4 F 0.0
15:00 221.5 F 0.0
20:00 220.1 F 0.0
25:00 221.9 F 0.0
30:00 219.0 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:46
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:11:11
CYCLE COMPLETE

04/18/2024 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____