

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4069

Name: RITTER 1  
Sterilizer ID: V2178198  
11/30/2023 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	71.8 F	0.0
2:00	74.5 F	0.2
4:00	84.3 F	1.0
6:00	97.9 F	2.7
8:00	128.5 F	4.9
10:00	177.3 F	7.1
12:00	184.0 F	10.4
14:00	192.0 F	13.5
16:00	204.5 F	3.9
18:00	214.2 F	6.9
20:00	223.6 F	10.2
22:00	233.1 F	10.0
24:00	235.1 F	9.6
26:00	240.2 F	10.9
28:00	243.5 F	12.2
30:00	254.1 F	17.6
32:00	263.5 F	23.2
34:00	272.7 F	29.1

STERILIZING

mm:ss	Degrees	PSI
0:00	271.3 F	28.4
1:00	272.4 F	29.1
2:00	271.4 F	28.4
3:00	272.1 F	29.2
4:00	272.0 F	28.8

Min 271.2 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	202.2 F	0.0
	5:00	210.2 F	0.0
	10:00	207.3 F	0.0
	15:00	213.1 F	0.0
	20:00	222.5 F	0.0
	25:00	232.0 F	0.0
Drying Complete	30:00	227.2 F	0.0

FILLING: 0:49  
HEATING: 34:25  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:12:58  
CYCLE COMPLETE

11/30/2023 10:09 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_