

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2410

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2022 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.2 F 0.0  
2:00 173.0 F 2.4  
4:00 189.2 F 6.7  
6:00 208.6 F 12.0  
8:00 226.3 F 7.0  
10:00 239.2 F 12.4  
12:00 237.6 F 9.5  
14:00 237.8 F 8.8  
16:00 248.9 F 14.5  
18:00 261.0 F 21.3  
20:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.6 F 29.2  
2:00 272.1 F 28.9  
3:00 271.9 F 28.5  
4:00 272.6 F 28.9

Min 270.8 F 27.9  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 231.4 F 0.0  
10:00 228.6 F 0.0  
15:00 227.9 F 0.0  
20:00 231.6 F 0.0  
25:00 228.1 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

10/06/2022 12:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_