
Midmark M11 - v1.0.5
Total Cycles: 1360

Name: RITTER2
Sterilizer ID: V2313228
08/03/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 78.3 F 0.0
4:00 110.4 F 0.0
6:00 138.5 F 0.0
8:00 164.3 F 1.9
10:00 186.8 F 5.3
12:00 204.2 F 9.0
14:00 221.6 F 12.8
16:00 224.0 F 6.4
18:00 231.6 F 10.3
20:00 237.4 F 9.1
22:00 239.3 F 11.4
24:00 236.7 F 9.3
26:00 249.5 F 15.3
28:00 260.1 F 21.4
30:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.5
2:00 271.3 F 28.8
3:00 272.3 F 29.6
4:00 271.5 F 28.9

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 215.2 F 0.0
10:00 216.8 F 0.0
15:00 234.1 F 0.0
20:00 237.6 F 0.0
25:00 227.8 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:57
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

08/03/2022 08:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____