
Midmark M11 - v1.0.5
Total Cycles: 356

Name: RITTER2
Sterilizer ID: V2313228
09/13/2021 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 155.0 F 3.0
4:00 182.1 F 7.3
6:00 197.3 F 12.4
8:00 224.0 F 7.0
10:00 232.8 F 12.3
12:00 236.8 F 10.0
14:00 236.7 F 8.9
16:00 249.7 F 15.5
18:00 261.6 F 22.3
20:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.4
2:00 271.2 F 28.6
3:00 272.3 F 29.7
4:00 272.1 F 29.1

Min 271.1 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 229.1 F 0.0
10:00 208.6 F 0.0
15:00 167.0 F 0.0
20:00 161.1 F 0.0
25:00 167.4 F 0.0
Dry Cycle Aborted

FILLING: 1:35
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:04
DRYING: 26:15
TOTAL CYCLE: 00:55:20

09/13/2021 12:51 PM
ACCEPTED: X
REJECTED:
Operator: U01