
Midmark M11 - v1.0.0
Total Cycles: 2908

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2023 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.8 F 0.0
2:00 159.8 F 3.0
4:00 182.3 F 6.6
6:00 195.6 F 10.8
8:00 215.7 F 5.0
10:00 222.6 F 8.7
12:00 234.0 F 13.4
14:00 235.4 F 9.8
16:00 235.2 F 8.8
18:00 245.7 F 13.4
20:00 257.4 F 19.6
22:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.3 F 29.3
2:00 271.4 F 28.9
3:00 272.3 F 29.4
4:00 271.4 F 28.8

Min 271.3 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 229.6 F 0.0
10:00 234.1 F 0.0
15:00 244.6 F 0.0
20:00 233.5 F 0.0
25:00 227.8 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

02/15/2023 02:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____