
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001809
13:27 26/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 1:59
STERILIZING 2:57
136.4 C 224kPa 2:57
136.4 C 225kPa 3:27
136.4 C 225kPa 3:57
136.3 C 223kPa 4:27
136.4 C 226kPa 4:57
136.4 C 223kPa 5:27
136.4 C 225kPa 5:57
136.2 C 224kPa 6:27
136.2 C 224kPa 6:27
VENTING 6:27
AIR DRYING 7:17
CYCLE COMPLETE 37:17

Digital Signature #
ED11B26DADB2AF52
