

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002306  
10:01 20/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:39  
PRESSURIZING 1:50  
STERILIZING 2:42  
136.0 C 220kPa 2:42  
136.1 C 224kPa 3:12  
136.4 C 226kPa 3:42  
136.4 C 226kPa 4:12  
136.1 C 222kPa 4:42  
136.1 C 221kPa 5:12  
136.4 C 224kPa 5:42  
136.2 C 223kPa 6:12  
136.2 C 223kPa 6:12  
VENTING 6:12  
AIR DRYING 7:05  
CYCLE COMPLETE 37:06  
-----

Digital Signature #  
02A32B4790E3EEA5  
-----