

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002492  
11:30 21/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:36  
STERILIZING 3:33  
136.0 C 220kPa 3:33  
136.4 C 222kPa 4:03  
136.4 C 225kPa 4:33  
136.4 C 225kPa 5:03  
136.4 C 221kPa 5:33  
136.4 C 224kPa 6:03  
136.1 C 218kPa 6:33  
136.1 C 220kPa 7:03  
136.1 C 220kPa 7:03  
VENTING 7:03  
AIR DRYING 7:56  
CYCLE COMPLETE 8:10  
-----

Digital Signature #  
9095231262567DD1  
-----