

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002455  
7:17 18/10/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:52  
STERILIZING 2:47  
136.0 C 220kPa 2:47  
136.3 C 220kPa 3:17  
136.4 C 220kPa 3:47  
136.3 C 222kPa 4:17  
136.3 C 221kPa 4:47  
136.4 C 220kPa 5:17  
136.2 C 218kPa 5:47  
136.3 C 220kPa 6:17  
136.3 C 220kPa 6:17  
VENTING 6:17  
AIR DRYING 7:10  
CYCLE COMPLETE 7:13  
-----

Digital Signature #  
67470545C9012F5A  
-----