

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1963

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2022 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
2:00 193.5 F 4.1  
4:00 206.3 F 8.9  
6:00 221.6 F 7.3  
8:00 231.5 F 9.5  
10:00 237.8 F 8.8  
12:00 243.3 F 12.8  
14:00 242.2 F 11.5  
16:00 255.4 F 18.3  
18:00 266.9 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.8 F 29.2  
2:00 272.0 F 28.6  
3:00 271.5 F 28.5  
4:00 272.7 F 29.3  
  
Min 271.4 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 215.7 F 0.0  
10:00 229.0 F 0.0  
15:00 229.5 F 0.0  
20:00 229.3 F 0.0  
25:00 230.7 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

05/27/2022 10:37 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_