
Midmark M11 - v1.0.0
Total Cycles: 4734

Name: RITTER 1
Sterilizer ID: V2178198
07/16/2024 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 175.6 F 3.1
4:00 181.0 F 7.6
6:00 201.3 F 12.8
8:00 220.1 F 7.4
10:00 231.5 F 12.7
12:00 234.6 F 9.7
14:00 236.7 F 8.9
16:00 246.9 F 14.2
18:00 258.9 F 20.7
20:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.9
2:00 272.2 F 29.3
3:00 271.4 F 28.8
4:00 272.1 F 29.7

Min 271.3 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 214.7 F 0.0
10:00 221.9 F 0.0
15:00 222.6 F 0.0
20:00 225.1 F 0.0
25:00 227.3 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

07/16/2024 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____