
Midmark M11 - v1.0.5
Total Cycles: 3785

Name: RITTER2
Sterilizer ID: V2313228
07/15/2024 05:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 173.3 F 3.7
4:00 195.8 F 9.1
6:00 225.8 F 7.5
8:00 235.0 F 10.7
10:00 232.7 F 8.3
12:00 239.2 F 10.3
14:00 247.6 F 14.5
16:00 260.9 F 21.8
18:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 273.0 F 29.9
2:00 272.7 F 30.0
3:00 272.3 F 30.1
4:00 271.9 F 29.5

Min 271.0 F 28.8
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 236.5 F 0.0
10:00 227.8 F 0.0
15:00 226.4 F 0.0
20:00 208.2 F 0.0
25:00 203.2 F 0.0
30:00 186.6 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

07/15/2024 06:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____