
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 004755
13:07 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:30
136.4 C 220kPa 3:30
136.3 C 224kPa 4:00
136.4 C 225kPa 4:30
136.1 C 223kPa 5:00
136.4 C 225kPa 5:30
136.4 C 222kPa 6:00
136.4 C 223kPa 6:30
136.4 C 223kPa 7:00
136.4 C 223kPa 7:00
VENTING 7:01
AIR DRYING 7:51
CYCLE COMPLETE 14:26

Digital Signature #
B06BFCECEA6F2650
