
Midmark M11 - v1.0.0
Total Cycles: 5090

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 170.1 F 0.2
4:00 183.2 F 0.7
6:00 215.0 F 2.0
8:00 213.4 F 6.2
10:00 227.1 F 10.4
12:00 233.3 F 7.8
14:00 239.2 F 11.4
16:00 237.1 F 9.1
18:00 239.6 F 9.7
20:00 248.9 F 14.3
22:00 260.5 F 20.7
24:00 270.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.9 F 29.0
2:00 272.4 F 28.4
3:00 272.0 F 28.6
4:00 272.7 F 28.8

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 222.2 F 0.0
10:00 218.4 F 0.0
15:00 199.6 F 0.0
20:00 199.1 F 0.0
25:00 221.6 F 0.0
30:00 251.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:36
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

11/14/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____