
Midmark M11 - v1.0.5
Total Cycles: 1921

Name: RITTER2
Sterilizer ID: V2313228
01/11/2023 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.6 F 0.0
2:00 183.8 F 4.1
4:00 195.5 F 9.1
6:00 220.6 F 7.0
8:00 228.0 F 9.6
10:00 233.6 F 8.6
12:00 244.2 F 13.4
14:00 244.2 F 12.7
16:00 257.6 F 19.9
18:00 269.3 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.6
2:00 271.7 F 28.9
3:00 271.6 F 29.4
4:00 272.3 F 29.6

Min 271.2 F 28.7
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 226.2 F 0.0
10:00 239.6 F 0.0
15:00 230.6 F 0.0
20:00 227.5 F 0.0
25:00 229.1 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

01/11/2023 03:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____