
Midmark M11 - v1.0.0
Total Cycles: 1724

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 190.8 F 3.7
4:00 211.0 F 8.4
6:00 228.3 F 9.0
8:00 233.9 F 10.0
10:00 233.3 F 7.7
12:00 245.4 F 12.6
14:00 245.2 F 12.7
16:00 258.4 F 19.8
18:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.5
2:00 272.2 F 29.0
3:00 271.7 F 28.3
4:00 272.2 F 29.0

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 242.1 F 0.0
10:00 214.3 F 0.0
15:00 220.1 F 0.0
20:00 225.9 F 0.0
25:00 224.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

03/16/2022 02:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____