
Midmark M11 - v1.0.0
Total Cycles: 3225

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 149.7 F 0.0
4:00 175.5 F 0.3
6:00 204.3 F 4.3
8:00 211.6 F 8.7
10:00 225.5 F 13.3
12:00 229.9 F 8.6
14:00 240.7 F 13.5
16:00 240.2 F 10.9
18:00 237.1 F 9.0
20:00 251.2 F 15.8
22:00 262.8 F 22.4
24:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.9
4:00 271.9 F 28.4

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 218.3 F 0.0
10:00 232.4 F 0.0
15:00 230.3 F 0.0
20:00 229.7 F 0.0
25:00 229.8 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

05/12/2023 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____