

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000639  
13:44 15/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:46  
STERILIZING 3:47  
136.0 C 220kPa 3:47  
136.3 C 224kPa 4:17  
136.4 C 225kPa 4:47  
136.4 C 225kPa 5:17  
136.4 C 222kPa 5:47  
136.3 C 223kPa 6:17  
136.1 C 223kPa 6:47  
136.4 C 225kPa 7:17  
136.4 C 225kPa 7:17  
VENTING 7:17  
AIR DRYING 8:09  
CYCLE COMPLETE 8:13

-----  
Digital Signature #  
5539B48D366756A5  
-----