
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001621
13:23 11/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:22
PRESSURIZING 6:03
STERILIZING 7:19
136.0 C 220kPa 7:19
136.1 C 221kPa 7:49
136.3 C 223kPa 8:19
136.4 C 222kPa 8:49
136.1 C 220kPa 9:19
136.3 C 223kPa 9:49
136.4 C 221kPa 10:19
136.4 C 224kPa 10:49
VENTING 10:49
AIR DRYING 11:53
CYCLE COMPLETE 14:32

Digital Signature #
B0D271A0F7F6671C
