
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008437
10:35 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:38
STERILIZING 4:40
136.0 C 220kPa 4:40
135.9 C 221kPa 5:10
136.3 C 221kPa 5:40
136.4 C 222kPa 6:10
136.4 C 225kPa 6:40
136.4 C 224kPa 7:10
136.3 C 223kPa 7:40
136.2 C 222kPa 8:10
136.2 C 222kPa 8:10
VENTING 8:10
AIR DRYING 9:06
CYCLE COMPLETE 11:25

Digital Signature #
45BE31306CF33414
