
Midmark M11 - v1.0.5
Total Cycles: 4174

Name: RITTER2
Sterilizer ID: V2313228
12/02/2024 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.4 F 0.0
2:00 86.9 F 0.0
4:00 97.5 F 0.2
6:00 131.2 F 0.8
8:00 157.6 F 3.3
10:00 184.0 F 6.6
12:00 200.3 F 10.2
14:00 217.1 F 7.5
16:00 227.0 F 8.0
18:00 235.4 F 12.5
20:00 236.0 F 9.6
22:00 238.6 F 9.6
24:00 246.5 F 13.8
26:00 257.9 F 20.0
28:00 268.2 F 26.4
30:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.6 F 29.3
2:00 273.0 F 29.6
3:00 272.2 F 29.7
4:00 272.1 F 29.1

Min 271.7 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 234.8 F 0.0
10:00 209.1 F 0.0
15:00 210.7 F 0.0
20:00 204.1 F 0.0
25:00 215.6 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 30:04
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

12/02/2024 05:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____