

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4469

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2025 08:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 73.5 F 0.0  
4:00 112.3 F 0.5  
6:00 110.6 F 2.4  
8:00 145.4 F 5.4  
10:00 178.9 F 8.8  
12:00 222.5 F 12.5  
14:00 214.8 F 4.0  
16:00 223.3 F 7.5  
18:00 233.8 F 11.3  
20:00 228.4 F 6.9  
22:00 241.3 F 12.1  
24:00 238.7 F 10.1  
26:00 251.0 F 16.1  
28:00 261.9 F 22.3  
30:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.7 F 29.4  
2:00 271.4 F 28.9  
3:00 272.8 F 29.3  
4:00 271.4 F 28.8

Min 271.2 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 182.1 F 0.0  
10:00 187.0 F 0.0  
15:00 188.7 F 0.0  
20:00 193.2 F 0.0  
25:00 201.6 F 0.0  
30:00 206.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 30:32  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:09:14  
CYCLE COMPLETE

03/28/2025 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_