
Midmark M11 - v1.0.0
Total Cycles: 2110

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2022 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.1 F 0.0
2:00 72.5 F 0.0
4:00 76.5 F 0.0
6:00 102.9 F 0.5
8:00 129.9 F 1.9
10:00 161.0 F 3.9
12:00 180.3 F 6.6
14:00 195.1 F 10.0
16:00 208.9 F 12.6
18:00 220.8 F 6.1
20:00 232.0 F 9.9
22:00 237.4 F 10.2
24:00 240.3 F 11.1
26:00 236.2 F 8.8
28:00 249.7 F 15.1
30:00 260.5 F 21.1
32:00 270.4 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.3 F 29.0
2:00 271.7 F 28.3
3:00 272.5 F 29.1
4:00 271.8 F 28.4

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 204.0 F 0.0
10:00 191.7 F 0.0
15:00 190.2 F 0.0
20:00 198.4 F 0.0
25:00 228.2 F 0.0
30:00 223.1 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:10:37
CYCLE COMPLETE

07/11/2022 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____