
Midmark M11 - v1.0.0
Total Cycles: 3337

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2023 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.0 F 0.0
2:00 195.0 F 1.4
4:00 220.4 F 6.9
6:00 232.7 F 12.5
8:00 232.5 F 8.6
10:00 241.8 F 10.8
12:00 244.0 F 12.6
14:00 243.0 F 11.7
16:00 256.5 F 18.7
18:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.7
2:00 271.9 F 28.6
3:00 272.4 F 28.6
4:00 271.7 F 28.5

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 220.1 F 0.0
10:00 222.6 F 0.0
15:00 226.9 F 0.0
20:00 227.4 F 0.0
25:00 226.9 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

06/05/2023 03:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____