
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001546
13:21 10/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:58
PRESSURIZING 6:36
STERILIZING 7:57
136.0 C 220kPa 7:57
136.4 C 223kPa 8:27
136.2 C 221kPa 8:57
136.2 C 219kPa 9:27
136.4 C 222kPa 9:57
136.4 C 221kPa 10:27
136.4 C 220kPa 10:57
136.2 C 221kPa 11:27
VENTING 11:27
AIR DRYING 12:29
CYCLE COMPLETE 13:22

Digital Signature #
20DD52ADBD149FF2
