
Midmark M11 - v1.0.5
Total Cycles: 3319

Name: RITTER2
Sterilizer ID: V2313228
01/24/2024 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 171.9 F 2.4
4:00 192.0 F 6.8
6:00 208.2 F 11.5
8:00 225.5 F 6.3
10:00 234.8 F 11.6
12:00 236.1 F 9.2
14:00 239.1 F 9.8
16:00 248.9 F 14.8
18:00 261.1 F 21.8
20:00 272.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.1
1:00 272.0 F 29.3
2:00 272.2 F 29.4
3:00 272.6 F 29.7
4:00 271.6 F 29.4

Min 271.5 F 29.0
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 226.6 F 0.0
10:00 192.6 F 0.0
15:00 188.4 F 0.0
20:00 184.6 F 0.0
25:00 182.5 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

01/24/2024 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____