
Midmark M11 - v1.0.0
Total Cycles: 2563

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2022 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 167.9 F 3.6
4:00 180.7 F 8.4
6:00 204.7 F 12.4
8:00 221.5 F 7.4
10:00 234.8 F 12.6
12:00 235.6 F 9.6
14:00 237.6 F 8.6
16:00 248.4 F 14.6
18:00 260.3 F 21.2
20:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.4
2:00 272.1 F 28.8
3:00 271.3 F 28.4
4:00 272.2 F 29.3

Min 271.3 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 211.7 F 0.0
10:00 220.9 F 0.0
15:00 226.2 F 0.0
20:00 228.4 F 0.0
25:00 229.5 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

11/15/2022 04:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____