

-----  
Midmark M11 - v1.0.0  
Total Cycles: 884

Name: RITTER 1  
Sterilizer ID: V2178198  
06/10/2021 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.8 F 0.0  
2:00 172.8 F 2.5  
4:00 192.2 F 6.6  
6:00 210.6 F 11.5  
8:00 221.2 F 5.7  
10:00 233.6 F 10.6  
12:00 233.5 F 7.7  
14:00 244.7 F 12.3  
16:00 245.7 F 12.9  
18:00 258.0 F 19.4  
20:00 268.9 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.4 F 28.4  
2:00 272.0 F 28.5  
3:00 272.1 F 28.3  
4:00 272.0 F 28.5  
  
Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 193.5 F 0.0  
10:00 206.1 F 0.0  
15:00 210.4 F 0.0  
20:00 242.3 F 0.0  
25:00 232.6 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 21:06  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

06/10/2021 03:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_