

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003242  
9:52 01/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:41  
PRESSURIZING 3:42  
STERILIZING 4:42  
136.4 C 220kPa 4:42  
136.4 C 223kPa 5:12  
135.9 C 219kPa 5:42  
136.2 C 219kPa 6:12  
136.2 C 221kPa 6:42  
136.2 C 218kPa 7:12  
136.4 C 222kPa 7:42  
136.4 C 223kPa 8:12  
136.4 C 223kPa 8:12  
VENTING 8:12  
AIR DRYING 9:07  
CYCLE COMPLETE 9:12

-----  
Digital Signature #  
E0A950AD298AF726  
-----