

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4185

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2024 05:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.5 F 0.0  
2:00 148.3 F 0.4  
4:00 175.0 F 2.7  
6:00 197.7 F 6.6  
8:00 211.2 F 11.1  
10:00 222.7 F 6.2  
12:00 232.4 F 11.4  
14:00 234.4 F 8.9  
16:00 239.6 F 10.5  
18:00 247.0 F 14.1  
20:00 258.6 F 20.5  
22:00 269.2 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 273.2 F 29.9  
2:00 272.5 F 30.1  
3:00 271.6 F 29.7  
4:00 272.1 F 29.6

Min 270.9 F 29.0  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 237.6 F 0.0  
10:00 186.6 F 0.0  
15:00 183.4 F 0.0  
20:00 180.9 F 0.0  
25:00 178.5 F 0.0  
30:00 177.9 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 23:05  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

12/04/2024 06:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_