
Midmark M11 - v1.0.5
Total Cycles: 874

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 06:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 72.6 F 0.0
4:00 79.5 F 0.0
6:00 107.0 F 0.2
8:00 140.7 F 2.4
10:00 165.9 F 5.4
12:00 217.1 F 8.9
14:00 212.6 F 12.8
16:00 216.0 F 6.6
18:00 227.5 F 11.0
20:00 232.1 F 7.2
22:00 242.2 F 12.5
24:00 240.8 F 10.8
26:00 252.6 F 16.9
28:00 263.4 F 23.1
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.8 F 29.5
2:00 271.7 F 29.5
3:00 272.1 F 29.3
4:00 272.1 F 29.8

Min 271.2 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 218.9 F 0.0
10:00 215.8 F 0.0
15:00 203.4 F 0.0
20:00 215.2 F 0.0
25:00 221.6 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:03
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:03
CYCLE COMPLETE

03/14/2022 08:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____