

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2334

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2023 03:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
2:00 203.4 F 4.3  
4:00 218.5 F 10.3  
6:00 225.2 F 6.4  
8:00 237.3 F 12.6  
10:00 238.9 F 11.0  
12:00 238.6 F 10.3  
14:00 253.5 F 17.6  
16:00 266.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.3 F 29.5  
2:00 271.2 F 28.9  
3:00 272.6 F 29.6  
4:00 271.7 F 29.0  
  
Min 271.2 F 28.7  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 222.4 F 0.0  
10:00 224.1 F 0.0  
15:00 229.8 F 0.0  
20:00 229.8 F 0.0  
25:00 227.0 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 2:39  
HEATING: 17:35  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:57:32  
CYCLE COMPLETE

05/03/2023 04:09 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_