
Midmark M11 - v1.0.0
Total Cycles: 2731

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 76.6 F 0.0
4:00 82.1 F 0.0
6:00 109.4 F 0.0
8:00 139.9 F 0.0
10:00 172.1 F 0.0
12:00 194.9 F 3.0
14:00 211.3 F 6.0
16:00 220.7 F 9.2
18:00 225.8 F 13.1
20:00 231.5 F 8.4
22:00 242.2 F 13.2
24:00 240.4 F 10.9
26:00 236.9 F 9.1
28:00 251.0 F 15.6
30:00 262.3 F 22.1
32:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 271.8 F 28.2
2:00 272.3 F 28.7
3:00 272.1 F 28.8
4:00 271.4 F 28.0

Min 271.3 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 209.1 F 0.0
10:00 213.3 F 0.0
15:00 199.3 F 0.0
20:00 200.2 F 0.0
25:00 215.6 F 0.0
30:00 222.9 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:19
CYCLE COMPLETE

01/04/2023 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____