
Midmark M11 - v1.0.0
Total Cycles: 1160

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 193.1 F 1.9
4:00 211.9 F 5.8
6:00 220.0 F 11.6
8:00 226.1 F 6.3
10:00 237.2 F 11.7
12:00 236.9 F 8.9
14:00 240.7 F 9.9
16:00 248.3 F 14.0
18:00 260.5 F 20.8
20:00 271.4 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.5 F 28.5
2:00 271.4 F 27.9
3:00 272.5 F 28.6
4:00 271.6 F 27.9

Min 271.4 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 233.1 F 0.0
10:00 199.8 F 0.0
15:00 229.0 F 0.0
20:00 201.6 F 0.0
25:00 200.3 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

09/13/2021 09:06 AM
ACCEPTED: X
REJECTED:
Operator: U01