
Midmark M11 - v1.0.0
Total Cycles: 2197

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2022 06:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 77.5 F 0.0
4:00 102.7 F 0.0
6:00 135.0 F 0.8
8:00 163.2 F 3.0
10:00 182.3 F 6.3
12:00 196.7 F 9.8
14:00 211.0 F 13.5
16:00 218.7 F 5.6
18:00 226.1 F 9.2
20:00 235.9 F 13.2
22:00 234.4 F 9.4
24:00 239.9 F 10.0
26:00 244.7 F 12.7
28:00 256.0 F 18.7
30:00 266.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.3 F 29.2
2:00 271.6 F 28.4
3:00 272.0 F 29.1
4:00 272.3 F 28.9

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 197.9 F 0.0
10:00 208.1 F 0.0
15:00 213.8 F 0.0
20:00 218.4 F 0.0
25:00 223.6 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

08/04/2022 07:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____