

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000348  
7:58 20/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:48  
STERILIZING 3:49  
136.0 C 229kPa 3:49  
136.4 C 232kPa 4:19  
136.0 C 230kPa 4:49  
136.1 C 231kPa 5:19  
136.2 C 231kPa 5:49  
136.1 C 229kPa 6:19  
136.4 C 233kPa 6:49  
136.3 C 230kPa 7:19  
136.3 C 230kPa 7:19  
VENTING 7:19  
AIR DRYING 8:11  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
957403C1E4E12D93  
-----